

Permanent Gas Analysis – Separation of Argon and Oxygen on a MolSieve 5A Column using the Agilent 490 Micro GC

Application Note

Micro Gas Chromatography, Permanent Gas Analysis

Authors

Mohamed Bajja and Remko van Loon
Agilent Technologies, Inc.
Middelburg
The Netherlands



Introduction

This application note shows an example of the analysis of permanent gases, including the separation of Argon and Oxygen, using the Agilent 490 Micro GC. For the separation of Argon and Oxygen, a High Resolution 20 m MolSieve 5A column is used.

The advantage of the Agilent 490 Micro GC is speed of analysis. Even with the 20 m HR MolSieve 5A column, you get the results fast. Total analysis time for the permanent gases until Nitrogen is approximately 3 minutes. The Agilent 490 Micro GC delivers lab-quality separations in an ultra-compact, portable instrument.




Agilent Technologies

Instrumentation

Instrument	Agilent 490 Micro GC (G3581A)
Column channel	20 m MolSieve 5A
Column temperature	40 °C
Carrier gas	Helium, 200 kPa
Injection time	40 msec

Sample information

Neon	18 ppm
Hydrogen	1.0 %
Argon	0.2 %
Oxygen	0.2 %
Nitrogen	0.2 %
Helium	matrix



For More Information

For more information on our products and services, visit our Web site at www.agilent.com/chem.

www.agilent.com/chem

Agilent shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc., 2011
Printed in the USA
July 27, 2011
5990-8700EN



Agilent Technologies